

STANDARD INFORMATION

Standard: UL 1069

Standard ID: Hospital Signaling and Nurse Call Equipment [UL 1069:2007 Ed.7+R:21Feb2022]

Previous Standard ID: Hospital Signaling and Nurse Call Equipment [UL 1069:2007 Ed.7+R:23Apr2020]

EFFECTIVE DATE OF NEW/REVISED REQUIREMENTS

Effective Date: **February 21, 2024**

IMPACT, OVERVIEW, AND ACTION REQUIRED

Impact Statement: Per our accreditation, Intertek is required to review reports against the standard revisions to confirm compliance. Once compliance is confirmed, the standard reference in the report is updated to show continued compliance to the technical requirements of the standard. Reports not updated to this version by the effective date above will be withdrawn.

Overview of Changes: New combined signal and receptacle assembly for nurse call signal system. Specific details of new/ revised requirements are found in table below.

Current Listings Not Active? – Please immediately identify any current Listing Reports or products that are no longer active and should be removed from our records. We will do this at no charge as long as Intertek is notified in writing prior to the review of your reports.



STANDARD INFORMATION

CLAUSE	VERDICT	COMMENT
	Info	COMBINED SIGNAL AND RECEPTACLE ASSEMBLY
		<i>New section added;</i>
		General
51		The following requirements are for combined signal and receptacle assembly for a nurse call signal system. These products are intended to be installed in an outlet box. See standard for details.
		<i>New section added;</i>
		Components
52		A Hospital Grade receptacle shall comply with the applicable requirements of the Standard for Attachment Plugs and Receptacles, UL 498. See standard for details.
		<i>New section added;</i>
		Construction
53		The assembly shall include hospital grade receptacle, screws, spacers, enclosure, electrical circuits/components for attachment to a wall outlet box. See standard for details.
		<i>New section added;</i>
		Enclosures
54		A faceplate for a duplex receptacle shall not prevent complete seating of a hospital grade attachment plug. See standard for details.
		<i>New section added;</i>
		Performance – Polymeric Materials Tests
55		When tested as described in 55.1.2, an insulating material used for an outlet box shall not ignite within 15 seconds after the application of the test current. See standard for details. See standard for details.
		<i>New section added;</i>
		Marking
56		A signaling unit shall be plainly and permanently marked where the marking is readily visible before or after installation, with the information required in Section 43. See standard for details.



CLAUSE	VERDICT	COMMENT
		<i>New section added;</i>
		Instructions and Installation Drawings
57		The assembly shall include the hospital grade receptacle, screws, spacers, enclosure, electrical circuits/components for attachment to a wall outlet box. See standard for details.
