

## STANDARD INFORMATION

**This SUN establishes the Continuing Certification approach to Electrical Equipment for Measurement, Control and Laboratory Use, Safety Requirements for Hand-Held Probe Assemblies for Electrical Measurement and Test**

**Standard Number:** UL 61010-031 / CSA C22.2 No. 61010-031

**Standard Name:** Electrical Equipment for Measurement, Control and Laboratory Use – Part 031: Safety Requirements for Hand-Held Probe Assemblies for Electrical Measurement and Test

**Standard Edition and Issue Date:** 2<sup>nd</sup> Edition Dated January 27, 2017

**Date of Revision:** January 27, 2017

**Date of Previous Revision of Standard:** UL 61010-031 1st Edition Revised November 7, 2014; CSA C22.2 No. 61010-031-07 Amended 2010: CSA C22.2 No. 61010-031A-07 (R2012).

## EFFECTIVE DATE OF NEW/REVISED REQUIREMENTS

**Effective Date:** **January 1, 2020**

## IMPACT, OVERVIEW, AND ACTION REQUIRED

**Impact Statement:** After January 1, 2020 existing certifications to UL 61010-031 / CSA C22/2 No. 61010-031 1<sup>st</sup> edition can remain at the 1<sup>st</sup> edition unless a change is made to the product design. **NOTE:** Effective immediately, this revised standard will be exclusively used for evaluation of new products unless the Applicant requests in writing that current requirements be used along with their understanding that their listings will be withdrawn on Effective Date noted above, unless the product is found to comply with new/revise requirements.

### Overview of Changes: Continued Certification Approach

After January 1, 2020

- All new listing must be evaluated to UL 61010-031 / CSA C22.2 No. 61010-031 2<sup>nd</sup> edition.
- Existing listings may remain at UL 61010-031 / CSA C22.2 No. 61010-031 1<sup>st</sup> edition unless:
  - A design change is made to the product that requires a certification decision or
  - A new/revise requirement in the future mat require action to be taken.

**If the applicable requirements noted in the table are not described in your report(s), these requirements will need to be confirmed as met and added to your report(s) such as markings, instructions, test results, etc. (as required).**

***Current Listings Not Active? – Please immediately identify any current Listing Reports or products that are no longer active and should be removed from our records. We will do this at no charge as long as Intertek is notified in writing prior to the review of your reports.***